


RefMat No



Included in  
COMAR Database:  
Green 1 --> YES  
Blue 0 --> NO

Category	<b>Critical Dimensions</b>
Category	<b>Lateral (X-Y-axis, 2-dim)</b>
Category	
Category	
Category	

Type\_of\_RefMat:

**Name of RM: NanoCD Standards (NCD)**

**Description:** NanoCD Standards (NCD);  
25, 45, 70, 110

**Link:** <http://www.vlsistandards.com/products/dimensional/nanocd.asp?sid=83>

Certified Quantities						
lateral distance	MeanValue	<b>25</b>	<b>nm</b>	Uncertainty	<input type="text"/>	<input type="text"/>
lateral distance	MeanValue	<b>45</b>	<b>nm</b>	Uncertainty	<input type="text"/>	<input type="text"/>
lateral distance	MeanValue	<b>70</b>	<b>nm</b>	Uncertainty	<input type="text"/>	<input type="text"/>
lateral distance	MeanValue	<b>110</b>	<b>nm</b>	Uncertainty	<input type="text"/>	<input type="text"/>
	MeanValue		<b>nm</b>	Uncertainty	<input type="text"/>	<input type="text"/>

Calibration of	<b>AFM</b>
Calibration of	
Calibration of	
Calibration of	
Calibration of	

Characterised by	<input type="text"/>
Characterised by	<input type="text"/>
Characterised by	<input type="text"/>
Characterised by	<input type="text"/>
Characterised by	<input type="text"/>

**2 Provider**

Provider No

Provider No  Web address

Provider

Contact Person	<input type="text" value="Marc Helvey, ext. 108"/>	Country	<input type="text" value="USA"/>
Email	<input type="text" value="sales.support@vlsistd.com"/>	City	<input type="text" value="San Jose"/>
Phone	<input type="text" value="(408) 428-1800"/>	City Code	<input type="text" value="CA 95134-2006"/>
		Street + No	<input type="text" value="3087 North First Street"/>