

RefMat No **34**


Included in  
COMAR Database:  
Green 1 --> YES  
Blue 0 --> NO

Category 1 **Lateral (X-Y-axis, 2-dim)**

Category 2

Category 3

Category 4

Category 5

Type\_of\_RefMat: **RM**

Name of RM: **2D100, 2D200**

Description **Lateral-(xy)-Calibration Standard  
2-dim lattice of inverted square pyramids,  
200nm pitch, etched into a silicon chip**

Link [http://www.nanosensors.com/Transfer\\_Standards.pdf](http://www.nanosensors.com/Transfer_Standards.pdf)

### Certified Quantities

<b>pitch; square pyramids</b>	MeanValue	<b>100</b>	<b>nm</b>	Uncertainty		
<b>pitch; square pyramids</b>	MeanValue	<b>200</b>	<b>nm</b>	Uncertainty		
	MeanValue			Uncertainty		
	MeanValue			Uncertainty		
	MeanValue			Uncertainty		

Calibration of

Calibration of

Calibration of

Calibration of

Calibration of

Characterised by

Characterised by

Characterised by

Characterised by

Characterised by

**2 Provider**

Provider No

**2**

Provider No **2** Web address <http://www.nanosensors.com>

Provider **NanoSensors**

Contact Person

Country

Switzerland

Email

info@nanosensors.com

City

Neuchatel

Phone

+41 (0)32 720 5085

City Code

CH-2002

Street + No

Rue Jaquet-Droz 1