

02.11.2011

RefMat No 57



Included in
COMAR Database:
Green 1 --> YES
Blue 0 --> NO

Category 1 Lateral (X-Y-axis, 1-dim)

Category 2

Category 3

Category 4

Category 5

Type_of_RefMat: CRM

Name of RM: Standard Microscale

Description Standard Microscale;
Calibration of length expansion scale on SEM

Link VLSI product ???

Certified Quantities

lateral distance	MeanValue	240	nm	Uncertainty	1	nm
	MeanValue		nm	Uncertainty		
	MeanValue		nm	Uncertainty		
	MeanValue		nm	Uncertainty		
	MeanValue		nm	Uncertainty		

Calibration of SEM

Calibration of

Calibration of

Calibration of

Calibration of

Characterised by

Characterised by

Characterised by

Characterised by

Characterised by

2 Provider

Provider No

17

Provider No 17 Web address <http://www.hitachi-hitec.com/global/index.html>

Provider Hitachi High-Technologies Co.Ltd.

Contact Person Country Japan

Email hht@nst.hitachi-hitec.com City Tokyo

Phone +81-3-3504-7111 City Code Minato-ku

Street + No 1-24-14 Nishi Shimbashi